



UK DECLARATION OF CONFORMITY

In accordance with UK Government Guidance

Manufacturer

Etac A/S
Parallelevæj 3, 8751 Gedved,
Denmark
Phone no. + 45 79685833
www.etac.com

UK Responsible Person

Etac Ltd
Unit 60, Hartlebury Trading Estate,
Hartlebury, Kidderminster, DY10
4JD
Email enquiries.uk@etac.com
Web www.etac.com/uk
Phone +44 (0)121 561 2222

Statement

This declaration of conformity is issued under the sole responsibility of the manufacturer.
The device(s) covered by this declaration of conformity is/are in conformity with EU
Regulation 2017/745 on medical devices.

Basic UDI-DI

570799515TL

Device description

Patient Slings

Intended Purpose

The Sling is an assistive device intended for alleviation of or compensation for a functional impairment due to an injury or disability. The device is designed for an individual lacking the ability to stand up and transfer oneself over shorter distances to another sitting position to/from a bed, a wheelchair, a chair, a toilet or similar, due to reduced mobility or physical strength.

Device name(s)

Molift RgoSling Toilet LowBack
Molift RgoSling Toilet High Back
Molift RgoSling Comfort HighBack
Molift RgoSling Shadow
Molift EvoSling HighBack Padded
Molift EvoSling HighBack Net Padded
Molift EvoSling LowBack Padded
Molift EvoSling LowBack Net Padded
Molift EvoSling Hygiene
Molift EvoSling Comfort MediumBack
Molift EvoSling Shadow
Molift UnoSling HighBack
Molift UnoSling Toilet
Molift RgoSling HighBack padded
Molift RgoSling HighBack Net
Molift RgoSling MediumBack Padded
Molift RgoSling MediumBack Net
Molift RgoSling MediumBack Plus
Molift EvoSling MediumBack Padded
Molift EvoSling MediumBack Net Padded

Parts covered in this declaration

Separate list available upon request.

Risk class of the device

Class I, Rule 1

Harmonized/Established Standards Separate list available upon request

Place Gedved, Denmark

Date of issue 11. November 2025

Name and function Sampavi Rajkumar, Quality Specialist



Signature, on behalf of Etac A/S